

ABX Minotrol CRP



LOT

MC 435

CONTROL



(Exp.) 2025-07-05
(YYYY-MM-DD)

Rev 2

PARAMETRES PARAMETERS		UNITES UNITS		ABX Lysebio													
				CONTROL			1	CONTROL			2	CONTROL			3	TOLERANCES TOLERANCE	
				MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE	MICROS		PENTRA	TOLERANCES TOLERANCE		
				EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP		EMI CRP	LC-767G	MS CRP			
GB	WBC	10 ³ /mm ³ ;10 ⁹ /l	2.0	2.0	1.9	± 0.4	7.7	7.6	7.9	± 1.0	20.6	20.5	21.3	± 2.2			
GR	RBC	10 ⁶ /mm ³ ;10 ¹² /l	2.30	2.28	2.24	± 0.15	4.63	4.56	4.66	± 0.18	5.48	5.45	5.53	± 0.25			
HB	HGB	g/dl	6.2	6.2	6.1	± 0.4	13.5	13.5	13.9	± 0.5	17.4	17.8	18.4	± 0.7			
		g/l	62	62	61	± 4	135	135	139	± 5	174	178	184	± 7			
		mmol/l	3.85	3.85	3.79	± 0.25	8.38	8.38	8.63	± 0.31	10.81	11.05	11.43	± 0.44			
HT	HCT	%	16.8	16.9	17.0	± 2.0	37.5	36.9	38.2	± 2.5	48.8	49.1	49.2	± 3.0			
		l/l	0.168	0.169	0.170	± 0.020	0.375	0.369	0.382	± 0.025	0.488	0.491	0.492	± 0.030			
VGM	MCV	µm ³ ; fl	73	74	76.0	± 4	81	81	82.0	± 4	89	90	89.0	± 4			
TGMH	MCH	pg	27.0	27.2	27.2	± 2.4	29.2	29.6	29.8	± 2.4	31.8	32.7	33.3	± 2.8			
		fmol	1.67	1.69	1.69	± 0.15	1.81	1.84	1.85	± 0.15	1.97	2.03	2.07	± 0.18			
CCMH	MCHC	g/dl	36.9	36.7	35.8	± 3.0	36.0	36.5	36.4	± 3.0	35.7	36.3	37.4	± 3.0			
		g/l	369	367	358	± 30	360	365	364	± 30	357	363	374	± 30			
		mmol/l	22.93	22.82	22.25	± 1.86	22.35	22.70	22.59	± 1.86	22.15	22.54	23.22	± 1.86			
IDR	RDW	%	14.0	14.0	15.5	± 3.0	14.0	13.0	13.0	± 3.0	14.0	13.0	11.5	± 3.0			
PLAQ.	PLTS	10 ³ /mm ³ ;10 ⁹ /l	69	70	67	± 20	247	244	243	± 40	452	477	485	± 60			
VPM	MPV	µm ³ ; fl	7.3	7.8	9.0	± 2.0	7.5	7.7	8.3	± 2.0	7.6	7.8	8.3	± 2.0			
LYMPHO		#	1.3	1.3	N/A	± 0.2	2.6	2.5	N/A	± 0.5	3.4	3.3	N/A	± 1.1			
		%	65.0	64.0	N/A	± 8.0	33.5	32.5	N/A	± 6.0	16.5	16.0	N/A	± 5.0			
MONO		#	0.2	0.2	N/A	± 0.2	0.6	0.5	N/A	± 0.5	2.1	1.5	N/A	± 1.4			
		%	9.5	9.5	N/A	± 8.6	7.5	7.0	N/A	± 6.3	10.0	7.5	N/A	± 6.8			
GRANULO		#	0.5	0.5	N/A	± 0.2	4.5	4.6	N/A	± 0.8	15.1	15.7	N/A	± 2.1			
		%	25.5	26.5	N/A	± 9.9	59.0	60.5	N/A	± 9.9	73.5	76.5	N/A	± 9.9			
CRP		mg/dl	0.46	0.51	0.52	± 0.40	2.49	2.64	2.79	± 0.80	4.20	4.23	4.72	± 1.50			
		mg/l	4.6	5.1	5.2	± 4.0	24.9	26.4	27.9	± 8.0	42.0	42.3	47.2	± 15.0			

Ref: TEMP-1070 Rev.43 BACK/VERSO 1300092170