

ABX Minotrol CRP



LOT

MC 434

CONTROL



(Exp.) 2025-04-05
(YYYY-MM-DD)

Rev 2

| PARAMETRES PARAMETERS | | UNITES UNITS | | ABX Lysebio | | | | | | | | | | | | | |
|--------------------------|------|-------------------------------------------------------|--------------|--------------|--------------|---------|-------------------------|--------------|--------------|---------|-------------------------|--------------|--------------|---------|-------------------------|-------------------------|--|
| | | | | CONTROL | | | 1 | CONTROL | | | 2 | CONTROL | | | 3 | TOLERANCES TOLERANCE | |
| | | | | MICROS | | PENTRA | TOLERANCES TOLERANCE | MICROS | | PENTRA | TOLERANCES TOLERANCE | MICROS | | PENTRA | TOLERANCES TOLERANCE | | |
| | | | | EMI CRP | LC-767G | MS CRP | | EMI CRP | LC-767G | MS CRP | | EMI CRP | LC-767G | MS CRP | | | |
| GB | WBC | 10 ³ /mm ³ ;10 ⁹ /l | 2.1 | 2.2 | 2.2 | ± 0.4 | 7.9 | 7.7 | 8.1 | ± 1.0 | 20.5 | 20.6 | 21.6 | ± 2.2 | | | |
| GR | RBC | 10 ⁶ /mm ³ ;10 ¹² /l | 2.29 | 2.27 | 2.26 | ± 0.15 | 4.63 | 4.58 | 4.64 | ± 0.18 | 5.52 | 5.55 | 5.60 | ± 0.25 | | | |
| HB | HGB | g/dl | 5.9 | 5.9 | 5.7 | ± 0.4 | 13.4 | 13.4 | 13.7 | ± 0.5 | 17.3 | 17.7 | 18.2 | ± 0.7 | | | |
| | | g/l | 59 | 59 | 57 | ± 4 | 134 | 134 | 137 | ± 5 | 173 | 177 | 182 | ± 7 | | | |
| | | mmol/l | 3.66 | 3.66 | 3.54 | ± 0.25 | 8.32 | 8.32 | 8.51 | ± 0.31 | 10.74 | 10.99 | 11.30 | ± 0.44 | | | |
| HT | HCT | % | 15.8 | 15.9 | 15.7 | ± 2.0 | 37.0 | 36.6 | 37.1 | ± 2.5 | 48.0 | 48.8 | 48.7 | ± 3.0 | | | |
| | | l/l | 0.158 | 0.159 | 0.157 | ± 0.020 | 0.370 | 0.366 | 0.371 | ± 0.025 | 0.480 | 0.488 | 0.487 | ± 0.030 | | | |
| VGM | MCV | µm ³ ; fl | 69 | 70 | 69.5 | ± 4 | 80 | 80 | 80.0 | ± 4 | 87 | 88 | 87.0 | ± 4 | | | |
| TGMH | MCH | pg | 25.8 | 26.0 | 25.2 | ± 2.4 | 28.9 | 29.3 | 29.5 | ± 2.4 | 31.3 | 31.9 | 32.5 | ± 2.8 | | | |
| | | fmol | 1.60 | 1.61 | 1.57 | ± 0.15 | 1.80 | 1.82 | 1.83 | ± 0.15 | 1.95 | 1.98 | 2.02 | ± 0.18 | | | |
| CCMH | MCHC | g/dl | 37.3 | 37.1 | 36.3 | ± 3.0 | 36.2 | 36.6 | 36.9 | ± 3.0 | 36.0 | 36.2 | 37.4 | ± 3.0 | | | |
| | | g/l | 373 | 371 | 363 | ± 30 | 362 | 366 | 369 | ± 30 | 360 | 362 | 374 | ± 30 | | | |
| | | mmol/l | 23.19 | 23.06 | 22.54 | ± 1.86 | 22.47 | 22.71 | 22.92 | ± 1.86 | 22.37 | 22.51 | 23.20 | ± 1.86 | | | |
| IDR | RDW | % | 14.5 | 14.0 | 15.0 | ± 3.0 | 14.5 | 13.5 | 13.0 | ± 3.0 | 14.5 | 12.5 | 12.5 | ± 3.0 | | | |
| PLAQ. | PLTS | 10 ³ /mm ³ ;10 ⁹ /l | 69 | 75 | 69 | ± 20 | 257 | 264 | 263 | ± 40 | 466 | 493 | 506 | ± 60 | | | |
| VPM | MPV | µm ³ ; fl | 7.4 | 8.4 | 8.8 | ± 2.0 | 8.1 | 8.2 | 9.0 | ± 2.0 | 7.7 | 7.9 | 8.1 | ± 2.0 | | | |
| LYMPHO | | # | 1.4 | 1.4 | N/A | ± 0.2 | 2.6 | 2.6 | N/A | ± 0.5 | 3.3 | 3.3 | N/A | ± 1.1 | | | |
| | | % | 68.0 | 65.5 | N/A | ± 8.0 | 33.0 | 33.5 | N/A | ± 6.0 | 16.0 | 16.0 | N/A | ± 5.0 | | | |
| MONO | | # | 0.1 | 0.2 | N/A | ± 0.1 | 0.6 | 0.6 | N/A | ± 0.5 | 1.4 | 1.3 | N/A | ± 1.2 | | | |
| | | % | 7.0 | 8.0 | N/A | ± 6.3 | 7.5 | 7.5 | N/A | ± 6.8 | 7.0 | 6.5 | N/A | ± 5.9 | | | |
| GRANULO | | # | 0.5 | 0.6 | N/A | ± 0.2 | 4.7 | 4.5 | N/A | ± 0.8 | 15.8 | 16.0 | N/A | ± 2.1 | | | |
| | | % | 25.0 | 26.5 | N/A | ± 9.9 | 59.5 | 59.0 | N/A | ± 9.9 | 77.0 | 77.5 | N/A | ± 9.9 | | | |
| CRP | | mg/dl | 0.51 | 0.54 | 0.47 | ± 0.40 | 2.56 | 2.69 | 2.44 | ± 0.80 | 4.39 | 4.50 | 4.59 | ± 1.50 | | | |
| | | mg/l | 5.1 | 5.4 | 4.7 | ± 4.0 | 25.6 | 26.9 | 24.4 | ± 8.0 | 43.9 | 45.0 | 45.9 | ± 15.0 | | | |

Ref: TEMP-1070 Rev.43 BACK/VERSO 1300092170