



LOT

MX 452 L
Rev 1

CONTROL L

(Exp.)

2025-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	2.1				2.1					± 0.4
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	2.11				2.10					± 0.17
HB HGB	g/dl	5.9				6.0					± 0.5
	g/l	59				60					± 5
	mmol/l	3.66				3.73					± 0.31
HT HCT	%	15.8				16.0					± 2.0
	l/l	0.158				0.160					± 0.020
VGM MCV	µm ³ ; fl	75				76					± 4
TGMH MCH	pg	28.0				28.6					± 2.0
	fmol	1.74				1.77					± 0.12
CCMH MCHC	g/dl	37.3				37.6					± 3.0
	g/l	373				376					± 30
	mmol/l	23.15				23.35					± 1.86
IDR RDW	%	15.5				14.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	78				76					± 22
VPM MPV	µm ³ ; fl	7.9				6.9					± 2.0
LYMPHO	#	1.32				1.31					± 0.40
	%	63.0				62.5					± 8.0
MONO	#	0.19				0.21					± 0.17
	%	9.0				10.0					± 8.1
GRANULO	#	0.59				0.58					± 0.40
	%	28.0				27.5					± 10.0

LOT

MX 452 N
Rev 1

CONTROL N

(Exp.)

2025-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse									TOLERANCES TOLERANCE
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA	PENTRA	
		ES60 Care ST				60 8P - 60 18P					
GB WBC	10 ³ /mm ³ ; 10 ⁹ /l	7.6				7.6					± 0.8
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	4.63				4.63					± 0.26
HB HGB	g/dl	13.6				13.7					± 0.7
	g/l	136				137					± 7
	mmol/l	8.45				8.51					± 0.43
HT HCT	%	38.0				37.5					± 2.7
	l/l	0.380				0.375					± 0.027
VGM MCV	µm ³ ; fl	82				81					± 4
TGMH MCH	pg	29.4				29.6					± 2.0
	fmol	1.82				1.84					± 0.12
CCMH MCHC	g/dl	35.8				36.5					± 3.0
	g/l	358				365					± 30
	mmol/l	22.25				22.69					± 1.86
IDR RDW	%	15.0				14.0					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	265				255					± 40
VPM MPV	µm ³ ; fl	7.6				6.8					± 2.0
LYMPHO	#	2.58				2.58					± 0.40
	%	34.0				34.0					± 5.0
MONO	#	0.57				0.61					± 0.51
	%	7.5				8.0					± 6.8
GRANULO	#	4.45				4.41					± 0.76
	%	58.5				58.0					± 10.0

Ref: TEMP-0830 Rev.43 FRONT / RECTO 1300104144



LOT

MX 452 H
Rev 1

CONTROL H

(Exp.)

2025-06-05
(YYYY-MM-DD)

PARAMETRES PARAMETERS	UNITES UNITS	ABX Lyse								TOLERANCES TOLERANCE	
		MICROS	MICROS	MICROS	MICROS	MICROS	ARGOS - HELIOS	PENTRA	PENTRA		PENTRA
		ES60 Care ST				60 8P - 60 18P					
GB WBC	10 ⁹ /mm ³ ; 10 ⁹ /l	20.8				20.7					± 1.7
GR RBC	10 ⁶ /mm ³ ; 10 ¹² /l	5.80				5.81					± 0.32
HB HGB	g/dl	18.2				18.3					± 0.9
	g/l	182				183					± 9
	mmol/l	11.30				11.36					± 0.56
HT HCT	%	51.6				51.7					± 3.4
	l/l	0.516				0.517					± 0.034
VGM MCV	µm ³ ; fl	89				89					± 4
TGMH MCH	pg	31.4				31.5					± 2.0
	fmol	1.95				1.96					± 0.12
CCMH MCHC	g/dl	35.3				35.4					± 3.0
	g/l	353				354					± 30
	mmol/l	21.90				21.98					± 1.86
IDR RDW	%	15.0				13.5					± 3.0
PLAQ. PLTS	10 ³ /mm ³ ; 10 ⁹ /l	540				524					± 65
VPM MPV	µm ³ ; fl	7.2				6.6					± 2.0
LYMPHO	#	3.33				3.31					± 1.00
	%	16.0				16.0					± 5.0
MONO	#	1.14				1.14					± 1.03
	%	5.5				5.5					± 5.0
GRANULO	#	16.33				16.25					± 2.07
	%	78.5				78.5					± 10.0